

| Analytes/Method | Level 1 - 55251 | Level 2 - 55252 | Units |
|---|-----------------|-----------------|-------|
| Bacteria | | | |
| Beckman Coulter IRIS Diagnostics iQ ²⁰⁰ Series Analyzers | May be present | Present | |
| KOVA [®] GLASSTIC [®] SLIDE 10 with GRIDS | May be present | Present | |
| Non-grid slides (~0.5 mL) | May be present | Present | |
| Non-grid slides (~1.0 mL) | May be present | Present | |
| Cen-Slide | May be present | Present | |
| Slide & Coverslip (~0.5 mL) | May be present | Present | |
| Slide & Coverslip (~1.0 mL) | May be present | Present | |
| FisherBrand UriSystem DeciSlide | May be present | Present | |

| Analytes | Level 1 - 55251 | Level 2 - 55252 |
|--|------------------------|-----------------------------|
| Accutest[®] URS 10 Urine Reagent Strips (VISUAL) | | |
| Leukocytes | Negative | 15 - 500 cells/µL (Tr - Lg) |
| Nitrite | Negative | Positive |
| Urobilinogen | Normal (0.2 - 1 mg/dL) | 2 - 8 mg/dL ⁷ |
| Protein | Negative | 30 - 2000 mg/dL (1+ - 4+) |
| pH | 5.0 - 6.0 | 7.0 - 8.5 |
| Blood | Negative - Trace | 25 - 200 cells/µL (Sm - Lg) |
| Specific Gravity | 1.005 - 1.020 | 1.000 - 1.015 |
| Ketones | Negative | 5 - 160 mg/dL (Tr - Lg) |
| Bilirubin | Negative | Small - Large |
| Glucose | Negative | 100 - 1000 mg/dL (Tr - 3+) |

| | | |
|---|--------------------|-----------------------------|
| Accutest[®] (Analyzers)⁹ | | |
| Glucose | Negative | 100 - 1000 mg/dL (Tr - 3+) |
| Bilirubin | Negative | Small - Large |
| Ketones | Negative | 5 - ≥80 mg/dL (Tr - Lg) |
| Specific Gravity | 1.010 - ≥1.030 | 1.000 - 1.015 |
| Blood | Negative - Trace | 25 - 200 cells/µL (Sm - Lg) |
| pH | 5.0 - 6.5 | 7.0 - 8.5 |
| Protein | Negative | Trace - 300 mg/dL |
| Urobilinogen | Normal (0.2 mg/dL) | 2 - 8 mg/dL ⁷ |
| Nitrite | Negative | Positive |
| Leukocytes | Negative | 15 - 500 cells/µL (Tr - Lg) |

| | | |
|--|----------------------|---|
| Accustrip[®] URS Reader/Visual | | |
| Bilirubin | Negative | 1 - 4 mg/dL (1+ - 3+) (17 - 70 µmol/L) ⁷ |
| Blood | Negative - 10 Ery/µL | 10 - 250 Ery/µL |
| Glucose | Negative - Normal | 50 - ≥500 mg/dL (8.3 - ≥27.8 mmol/L) |
| Ketones | Negative | 25 - 300 mg/dL (1+ - 3+) (2.5 - 30 mmol/L) |
| Leukocytes | Negative | 25 - 500 Leu/µL |
| Nitrite | Negative | Positive |
| pH | 5.0 - 7.0 | 7.0 - 9.0 |
| Protein | Negative | 30 - 500 mg/dL (0.3 - 5.0 g/L) (1+ - 3+) |
| Specific Gravity (Density) | 1.010 - 1.025 | 1.000 - 1.020 |
| Urobilinogen | Normal | 2 - 12 mg/dL (34 - 200 µmol/L) ⁷ |

| | | |
|---|------------------|---------------------------|
| Beckman Coulter IRIS Diagnostics[®] iChem[®] VELOCITY[™] Analyzer | | |
| Bilirubin | Negative | Not Compatible |
| Urobilinogen | Normal | 2 - 4 mg/dL (1+ - 2+) |
| Ketones | Negative | 5 - 80 mg/dL (Tr+ - 2+) |
| Ascorbic Acid | Negative | Negative |
| Glucose | Negative | 50 - 500 mg/dL (1+ - 3+) |
| Protein | Negative | 30 - 500 mg/dL (1+ - 3+) |
| Blood | Negative - Trace | 0.03 - ≥1 mg/dL (1+ - 3+) |
| pH | 5.0 - 7.0 | 7.0 - 9.0 |
| Nitrite | Negative | Positive |
| Leukocytes | Negative | 75 - 500 mg/dL (2+ - 3+) |
| Specific Gravity | 1.020 - 1.026 | 1.012 - 1.018 |

| | | |
|-------------------------------------|------------------------|---------------------------------|
| Confirmatory and Other Tests | | |
| K-CHECK (Ketones) | Negative | Small - Large |
| Ictotest (Bilirubin) | Negative | Positive |
| Refractometer (Specific Gravity) | 1.021 - 1.027 | 1.013 - 1.019 |
| hCG | Negative | Positive |
| pH Paper | 4.0 - 6.0 | 7.0 - 9.0 |
| Sulfosalicylic Acid (Total Protein) | (≤ 0.05) ¹⁰ | Positive (≥ 0.50) ¹⁰ |

| Analytes | Level 1 - 55251 | Level 2 - 55252 |
|--|----------------------|---|
| Henry Schein One Step Plus/Urispec Plus Analyzer (Visual) | | |
| Bilirubin | Negative | 1 - 4 mg/dL (1+ - 3+) (17 - 70 µmol/L) ⁴ |
| Blood | Negative - 10 Ery/µL | 10 - 250 Ery/µL |
| Glucose | Negative - Normal | 50 - ≥500 mg/dL (8.3 - ≥27.8 mmol/L) ⁸ |
| Ketones | Negative | 25 - 300 mg/dL (1+ - 3+) (2.5 - 30 mmol/L) |
| Leukocytes | Negative | 25 - 500 Leu/µL |
| Nitrite | Negative | Positive |
| pH | 5.0 - 7.0 | 7.0 - 9.0 |
| Protein | Negative | 30 - 500 mg/dL (0.3 - 5.0 g/L) (1+ - 3+) |
| Specific Gravity (Density) | 1.005 - 1.025 | 1.000 - 1.020 |
| Urobilinogen | Normal | 2 - 12 mg/dL (34 - 200 µmol/L) ⁴ |

| | | |
|---|----------------------|-----------------------------|
| Henry Schein Urispec[®] 10SG (Visual) | | |
| Leukocytes | Negative | 15 - 500 Cells/µL (Tr - Lg) |
| Nitrite | Negative | Positive |
| Urobilinogen | Normal (0.2 E.U./dL) | 1 - 8 E.U./dL |
| Protein | Negative | Trace - 300 mg/dL (Tr - 3+) |
| pH | 5.0 - 6.5 | 6.5 - 8.5 |
| Blood | Negative - 10 Ery/µL | 10 - 200 Cells/µL (Tr - Lg) |
| Specific Gravity (1.0-) | 1.015 - 1.030 | 1.000 - 1.015 |
| Ketones | Negative | 5 - 160 mg/dL (Tr - Lg) |
| Bilirubin | Negative | Small - Large (1+ - 3+) |
| Glucose | Negative | 100 - 1000 mg/dL (Tr - 3+) |

| | | |
|---|-------------------|---|
| MACHEREY-NAGEL[®] URYXXON[®] Relax/300/500 Analyzer / VISUAL | | |
| Bilirubin | Negative | 1 - 4 mg/dL (1+ - 3+) (17 - 70 µmol/L) ⁷ |
| Blood | Negative - Trace | 10 - 250 Ery/µL |
| Glucose | Negative - Normal | 50 - ≥500 mg/dL (8.3 - ≥27.8 mmol/L) ⁸ |
| Ketones | Negative | 25 - 300 mg/dL (1+ - 3+) (2.5 - 30 mmol/L) |
| Leukocytes | Negative | 25 - 500 Leu/µL |
| Nitrite | Negative | Positive |
| pH | 5.0 - 7.0 | 7.0 - 9.0 |
| Protein | Negative | 30 - 500 mg/dL (0.3 - 5.0 g/L) (1+ - 3+) |
| Specific Gravity | 1.005 - 1.025 | 1.000 - 1.020 |
| Urobilinogen | Normal | 2 - 12 mg/dL (34 - 200 µmol/L) ⁷ |

| | | |
|--|----------------------|---------------------------|
| McKesson[®] (Consult Diagnostics 10SG Urine Reagent Strips) (Visual) | | |
| Glucose | Negative | 100 - 1000 mg/dL (± - 3+) |
| Bilirubin | Negative | 1 - 4 mg/dL (1+ - 3+) |
| Ketones | Negative | 5 - 160 mg/dL (± - 4+) |
| Specific Gravity | 1.015 - 1.030 | 1.000 - 1.020 |
| Blood | Negative - Trace | 1+ - 3+ |
| pH | 5.0 - 6.5 | 7.0 - 9.0 |
| Protein | Negative | 15 - 300 mg/dL (± - 3+) |
| Urobilinogen | Normal (0.2 E.U./dL) | 1 - 8 mg/dL |
| Nitrite | Negative | Positive |
| Leukocytes | Negative | 15 - 500 Leu/µL (± - 3+) |

| | | |
|--|----------------------|-------------------------------|
| McKesson[®] 120 Urine Analyzer | | |
| Leukocytes | Negative | 15 - 500 Leu/µL (± - 3+) |
| Nitrite | Negative | Positive |
| Urobilinogen | Normal (0.2 E.U./dL) | 0.2 - 2.0 mg/dL ¹¹ |
| Protein | Negative | Trace - 100 mg/dL (± - 2+) |
| pH | 5.0 - 6.5 | 7.0 - 9.0 |
| Blood | Negative - Trace | 25 - 200 Ery/µL (1+ - 3+) |
| Specific Gravity | 1.015 - 1.030 | 1.000 - 1.020 |
| Ketones | Negative | 5 - 80 mg/dL (Tr - 3+) |
| Bilirubin | Negative | 1 - 4 mg/dL (1+ - 3+) |
| Glucose | Negative | 100 - ≥1000 mg/dL (Tr - 3+) |

| | | |
|---|--------------------|-------------------------|
| ROCHE VISUAL TESTING (Visual Test Strips Only) | | |
| Specific Gravity | 1.015 - 1.030 | 1.000 - 1.010 |
| pH | 5 - 6 | 7 - 9 |
| Leukocytes | Negative | 1+ - 2+ |
| Nitrite | Negative | Positive |
| Protein | Negative | 30 - 100 mg/dL |
| Glucose | Normal | 250 - 1000 mg/dL |
| Ketones | Negative | Small - Large (1+ - 3+) |
| Urobilinogen ^{9*} | Normal | 1 - 8 mg/dL (1+ - 3+) |
| Bilirubin ^{9*} | Negative | 1+ - 3+ |
| Blood | Negative - Trace | 50 - 250 Ery/µL |
| Microalbumin ⁶ | Negative - 20 mg/L | 50 - 100 mg/L |

| Analytes | Level 1 - 55251 | Level 2 - 55252 |
|---|-----------------------|--|
| ROCHE Chemstrip 101 Urine Analyzer or ROCHE Urisys 1100 Urine Analyzer | | |
| Blood | Negative - Trace | 50 - 250 Ery/ μ L (1+ - 2+) |
| Bilirubin | Negative | 3 - 6 mg/dL (2+ - 3+) * |
| Urobilinogen | Normal | 1 - 8 mg/dL (1+ - 3+) ⁷ * |
| Ketones | Negative | 15 - 150 mg/dL (1+ - 3+) |
| Glucose | Normal | 250 - 1000 mg/dL (2+ - 3+) |
| Protein | Negative ⁶ | 15 (Tr) - 100 mg/dL (Tr - 2+) ⁷ |
| Nitrite | Negative | Positive |
| Leukocytes | Negative | 75 - 500 Leu/ μ L (1+ - 2+) |
| pH | 5 - 6.5 | 7 - 9 |
| Specific Gravity | 1.015 - 1.025 | 1.000 - 1.015 |
| ROCHE Chemstrip UA Urine Analyzer | | |
| Blood | Negative - Trace | 150 - 250 Ery/ μ L |
| Bilirubin | Negative | 3 - 6 mg/dL (2+ - 3+) * |
| Urobilinogen | Normal | 1 - 8 mg/dL ⁷ * |
| Ketones | Negative | 15 - 150 mg/dL (1+ - 3+) |
| Glucose | Normal | 250 - 1000 mg/dL |
| Protein | Negative | 30 - 100 mg/dL (1+ - 2+) ⁷ |
| Nitrite | Negative | Positive |
| Leukocytes | Negative | 100 - 500 Leu/ μ L (1+ - 2+) |
| pH | 5 - 6.5 | 7 - 9 |
| Specific Gravity | 1.015 - 1.025 | 1.000 - 1.015 |
| ROCHE Cobas 6500 (cobas u 601) | | |
| Blood | Negative - Trace | 150 - 250 Ery/ μ L |
| Leukocytes | Negative | 100 - 500 Leu/ μ L |
| Nitrite | Negative | Positive |
| Ketones | Negative | 15 - 150 mg/dL |
| Glucose | Normal | 250 - 1000 mg/dL |
| Protein | Negative ⁶ | 30 - 100 mg/dL ⁷ |
| Urobilinogen | Normal | 1 - 8 mg/dL ⁷ * |
| Bilirubin | Negative | 3 - 6 mg/dL * |
| pH | 5 - 6 | 7 - 9 |
| Specific Gravity | 1.019 - 1.030 | 1.011 - 1.021 |
| ROCHE cobas u 411 | | |
| Blood | Negative - Trace | 50 - 250 Ery/ μ L (3+ - 5+) |
| Bilirubin | Negative | 3 - 6 mg/dL (2+ - 3+) * |
| Urobilinogen | Negative - Normal | 1 - 8 mg/dL (1+ - 3+) ⁷ * |
| Ketones | Negative | 15 - 150 mg/dL (1+ - 4+) |
| Glucose | Negative - Normal | 250 - 1000 mg/dL (3+ - 4+) |
| Protein | Negative ⁶ | 30 - 100 mg/dL (2+ - 3+) ⁷ |
| Nitrite | Negative | Positive |
| Leukocytes | Negative | 100 - 500 Leu/ μ L (2+ - 3+) |
| pH | 5 - 6.5 | 7 - 9 |
| Specific Gravity | 1.015 - 1.025 | 1.000 - 1.020 |
| ROCHE Urisys 1800 | | |
| Blood | Negative - Trace | 25 - 250 Ery/ μ L (2+ - 5+) |
| Bilirubin | Negative | 3 - 6 mg/dL (2+ - 3+) * |
| Urobilinogen | Negative - Normal | 1 - 8 mg/dL (1+ - 3+) ⁷ * |
| Ketones | Negative | 15 - 150 mg/dL (1+ - 4+) |
| Glucose | Negative - Normal | 300 - 1000 mg/dL (3+ - 4+) |
| Protein | Negative ⁶ | 75 - 500 mg/dL (2+ - 4+) ⁷ |
| Nitrite | Negative | Positive |
| Leukocytes | Negative | 100 - 500 Leu/ μ L (2+ - 3+) |
| pH | 5 - 6.5 | 7 - 9 |
| Specific Gravity | 1.015 - 1.025 | 1.000 - 1.020 |
| ROCHE Urisys 2400 | | |
| Blood | Negative - Trace | 25 - 250 Ery/ μ L |
| Bilirubin | Negative | 3 - 6 mg/dL * |
| Urobilinogen | Normal | 1 - 8 mg/dL ⁸ * |
| Ketones | Negative | 15 - 150 mg/dL |
| Glucose | Normal | 300 - 1000 mg/dL |
| Protein | Negative ⁶ | 25 - 150 mg/dL ⁸ |
| Nitrite | Negative | Positive |
| Leukocytes | Negative | 100 - 500 Leu/ μ L |
| pH | 5 - 6.5 | 7 - 9 |
| Specific Gravity | 1.018 - 1.029 | 1.010 - 1.021 |

| Analytes | Level 1 - 55251 | Level 2 - 55252 |
|---|-----------------------------------|---------------------------------------|
| SIEMENS VISUAL TESTING (Visual Test Strips Only) | | |
| Glucose | Negative | 100 - 1000 mg/dL |
| Bilirubin | Negative | Small - Large (1+ - 3+) |
| Ketones | Negative | Trace(5) - 160 mg/dL (Tr - Lg) |
| Specific Gravity | 1.015 - 1.030 | 1.000 - 1.015 |
| Blood | Negative - Trace | Trace - Large (Tr - 3+) |
| pH | 5 - 6.5 | 7.0 - 8.5 |
| Protein | Negative | 30 - 300 mg/dL (1+ - 3+) |
| Urobilinogen | Normal (0.2 E.U./dL) ⁶ | 1.0 - 8.0 E.U./dL * |
| Nitrite | Negative | Positive |
| Leukocytes | Negative | Trace - Large (Tr - 3+) |
| Creatinine ² | 10 - 100 mg/dL | 100 - 300 mg/dL |
| SIEMENS[®] CLINITEK 50 | | |
| Glucose | Negative | 100 - 500 mg/dL (Tr - 2+) |
| Bilirubin | Negative | Small - Large (1+ - 3+) |
| Ketones | Negative | Trace - \geq 80 mg/dL (\pm - 3+) |
| Specific Gravity | 1.010 - \geq 1.030 | \leq 1.005 - 1.020 |
| Blood | Negative - Trace | Small - Large (1+ - 3+) |
| pH | 5.0 - 6.5 | 7 - \geq 9 |
| Protein | Negative | 30 - \geq 300 mg/dL (1+ - 3+) |
| Urobilinogen | Normal (0.2 E.U./dL) | 1.0 - 4.0 E.U./dL |
| Nitrite | Negative | Positive |
| Leukocytes | Negative | Trace - Large (Tr - 3+) |
| Microalbumin ¹ | \leq 10 - 30 mg/L | 30 - 300 mg/L |
| Creatinine ³ | 10 - 100 mg/dL | 100 - 300 mg/dL |
| SIEMENS[®] CLINITEK 500 | | |
| Glucose | Negative | 100 - 500 mg/dL (Tr - 2+) |
| Bilirubin | Negative | Small - Large (1+ - 3+) |
| Ketones | Negative | Trace - \geq 80 mg/dL (\pm - 3+) |
| Specific Gravity | 1.015 - \geq 1.030 | \leq 1.005 - 1.020 |
| Blood | Negative - Trace | Small - Large (1+ - 3+) |
| pH | 5 - 6.5 | 7.5 - \geq 9.0 |
| Protein | Negative | 30 - \geq 300 mg/dL (1+ - 3+) |
| Urobilinogen | Normal (0.2 E.U./dL) | 1.0 - \geq 8.0 E.U./dL |
| Nitrite | Negative | Positive |
| Leukocytes | Negative | Trace - Large (Tr - 3+) |
| Creatinine ² | 10 - 100 mg/dL | 100 - 300 mg/dL |
| SIEMENS[®] CLINITEK ADVANTUS | | |
| Glucose | Negative | 100 - 500 mg/dL (Tr - 2+) |
| Bilirubin | Negative | Small - Large (1+ - 3+) |
| Ketones | Negative | Trace - \geq 80 mg/dL (Tr - 3+) |
| Specific Gravity | 1.015 - \geq 1.030 | \leq 1.005 - 1.020 |
| Blood | Negative - Trace | Small - Large (1+ - 3+) |
| pH | 5.0- 6.5 | 7.0 - \geq 9.0 |
| Protein | Negative | 30 - \geq 300 mg/dL (1+ - 3+) |
| Urobilinogen | Normal (0.2 E.U./dL) | 2.0 - \geq 8.0 E.U./dL |
| Nitrite | Negative | Positive |
| Leukocytes | Negative | Trace - Large (Tr - 3+) |
| Creatinine ² | 10 - 100 mg/dL | 100 - 300 mg/dL |
| SIEMENS CLINITEK STATUS/ STATUS PLUS/STATUS CONNECT | | |
| Glucose | Negative | 100 - 500 mg/dL (Tr - 2+) |
| Bilirubin | Negative | Small - Large (1+ - 3+) |
| Ketones | Negative | Trace - \geq 160 mg/dL (Tr - 4+) |
| Specific Gravity | 1.010 - \geq 1.030 | \leq 1.005 - 1.025 |
| Blood | Negative - Trace | Small - Large (1+ - 3+) |
| pH | 5.0 - 6.5 | 7 - \geq 9 |
| Protein | Negative | 30 - \geq 300 mg/dL (1+ - 3+) |
| Urobilinogen | Normal (0.2 E.U./dL) | 2.0 - \geq 8.0 E.U./dL |
| Nitrite | Negative | Positive |
| Leukocytes | Negative | Trace - Large (Tr - 3+) |
| Microalbumin ¹ | \leq 10 - 30 mg/L | 30 - 300 mg/L |
| Creatinine ³ | 10 - 100 mg/dL | 100 - 300 mg/dL |
| hCG | Negative | Positive |
| Teco Diagnostics TC-101 • TC-201 • TC-720 Urine Analyzer | | |

DATA NOT AVAILABLE AT THE TIME OF PRINTING.

| Analytes | Level 1 - 55251 | Level 2 - 55252 |
|--|-------------------|---------------------------------------|
| Uriscan™ • 10 SGL Strips (Visual) | | |
| Blood | Negative - Trace | 10 - 250 RBC/µL (1+ - 3+) |
| Bilirubin | Negative | 0.5 - 3.0 mg/dL (1+ - 3+) |
| Urobilinogen | Negative - Normal | 1 - 12 mg/dL (1+ - 4+) |
| Ketones | Negative | 5 - 100 mg/dL (± - 3+) |
| Protein | Negative | 30 - 300 mg/dL (1+ - 3+) ⁷ |
| Nitrite | Negative | Positive |
| Glucose | Negative | 100 - 1000 mg/dL (± - 3+) |
| pH | 5.0 - 6.0 | 7.0 - 9.0 |
| Specific Gravity | 1.015 - 1.030 | 1.000 - 1.015 |
| Leukocytes | Negative | 25 - 500 WBC/µL (1+ - 3+) |
| Uriscan™ Optima Urine Analyzers • 10 SGL Strips | | |
| Blood | Negative - Trace | 10 - 250 RBC/µL (1+ - 3+) |
| Bilirubin | Negative | 0.5 - 3.0 mg/dL (1+ - 3+) |
| Urobilinogen | Negative - Normal | 1 - 12 mg/dL (1+ - 4+) |
| Ketones | Negative | 5 - 100 mg/dL (± - 3+) |
| Protein | Negative | 10 - 300 mg/dL (± - 3+) ⁷ |
| Nitrite | Negative | Positive |
| Glucose | Negative | 100 - 1000 mg/dL (± - 3+) |
| pH | 5.0 - 6.5 | 6.5 - 8.0 |
| Specific Gravity | 1.015 - 1.030 | 1.005 - 1.020 |
| Leukocytes | Negative | 25 - 500 WBC/µL (1+ - 3+) |

INTERNATIONAL USE ONLY

This Section is for International Use only and contains data for methods that are not available or cleared for diagnostic use in the United States.

| Method | Level 1 - 55251 | Level 2 - 55252 | Units |
|--|---|-----------------|-------|
| Red Blood Cells (Erythrocytes) | | | |
| 77 Elektronika UriSed Variants / COBIO Variants | 0 - 30 | 0 - 150 | p/µL |
| Beckman Coulter IRIS Diagnostics iQ@200 Series Analyzers | 0 - 10 | 8 - 43 | /µL |
| Mindray EH-2030, EH-2050-A Plus, EH-2050-B Plus, EH-2080B, EH-2080-C | DATA NOT AVAILABLE AT THE TIME OF PRINTING. | | |
| YD Diagnostics URISCAN PLUSCOPE | 0 - 50 | 0 - 200 | p/µL |
| ROCHE cobas 6500 (cobas u 701) | 0 - 50 | 22 - 69 | p/µL |
| KOVA® GLASSTIC® SLIDE 10 with GRIDS | 3 - 13 | 18 - 95 | /µL |
| Non-grid slides (~0.5 mL) | 0 - 5 | 4 - 38 | /hpf |
| Non-grid slides (~1.0 mL) | 0 - 4 | 1 - 25 | /hpf |
| Cen-Slide | 0 - 5 | 3 - 40 | /hpf |
| Slide & Coverslip (~0.5 mL) | 0 - 6 | 3 - 21 | /hpf |
| Slide & Coverslip (~1.0 mL) | 0 - 4 | 1 - 13 | /hpf |
| FisherBrand UriSystem DeciSlide | 0 - 5 | 4 - 31 | /hpf |
| White Blood Cells (Leukocytes) | | | |
| 77 Elektronika UriSed Variants / COBIO Variants | 0 - 30 | 20 - 90 | p/µL |
| Beckman Coulter IRIS Diagnostics iQ@200 Series Analyzers | 2 - 15 | 25 - 71 | /µL |
| Mindray EH-2030, EH-2050-A Plus, EH-2050-B Plus, EH-2080B, EH-2080-C | DATA NOT AVAILABLE AT THE TIME OF PRINTING. | | |
| YD Diagnostics URISCAN PLUSCOPE | 0 - 50 | 0 - 200 | p/µL |
| ROCHE cobas 6500 (cobas u 701) | 0 - 25 | 35 - 106 | p/µL |
| KOVA® GLASSTIC® SLIDE 10 with GRIDS | 2 - 15 | 17 - 85 | /µL |
| Non-grid slides (~0.5 mL) | 0 - 9 | 1 - 34 | /hpf |
| Non-grid slides (~1.0 mL) | 0 - 4 | 1 - 19 | /hpf |
| Cen-Slide | 0 - 6 | 0 - 41 | /hpf |
| Slide & Coverslip (~0.5 mL) | 0 - 4 | 1 - 19 | /hpf |
| Slide & Coverslip (~1.0 mL) | 0 - 4 | 0 - 10 | /hpf |
| FisherBrand UriSystem DeciSlide | 0 - 5 | 2 - 22 | /hpf |
| Casts | | | |
| 77 Elektronika UriSed Variants / COBIO Variants | None | None | |
| Mindray EH-2030, EH-2050-A Plus, EH-2050-B Plus, EH-2080B, EH-2080-C | DATA NOT AVAILABLE AT THE TIME OF PRINTING. | | |
| YD Diagnostics URISCAN PLUSCOPE | None | None | |
| KOVA® GLASSTIC® SLIDE 10 with GRIDS | None | None | |
| Non-grid slides (~0.5 mL) | None | None | |
| Non-grid slides (~1.0 mL) | None | None | |
| Cen-Slide | None | None | |
| Slide & Coverslip (~0.5 mL) | None | None | |
| Slide & Coverslip (~1.0 mL) | None | None | |
| FisherBrand UriSystem DeciSlide | None | None | |

| Method | Level 1 - 55251 | Level 2 - 55252 | Units |
|--|---|-----------------|-------|
| Crystals | | | |
| 77 Elektronika UriSed Variants / COBIO Variants | may be present | present | |
| Beckman Coulter IRIS Diagnostics iQ@200 Series Analyzers | none | present | |
| Mindray EH-2030, EH-2050-A Plus, EH-2050-B Plus, EH-2080B, EH-2080-C | DATA NOT AVAILABLE AT THE TIME OF PRINTING. | | |
| YD Diagnostics URISCAN PLUSCOPE | none | may be present | |
| ROCHE cobas 6500 (cobas u 701) | negative | positive | |
| KOVA® GLASSTIC® SLIDE 10 with GRIDS | none | present | |
| Non-grid slides (~0.5 mL) | none | present | |
| Non-grid slides (~1.0 mL) | none | present | |
| Cen-Slide | none | present | |
| Slide & Coverslip (~0.5 mL) | none | present | |
| Slide & Coverslip (~1.0 mL) | none | present | |
| FisherBrand UriSystem DeciSlide | none | present | |
| Bacteria | | | |
| 77 Elektronika UriSed Variants / COBIO Variants | may be present | present | |
| Beckman Coulter IRIS Diagnostics iQ@200 Series Analyzers | may be present | may be present | |
| KOVA® GLASSTIC® SLIDE 10 with GRIDS | may be present | present | |
| Non-grid slides (~0.5 mL) | may be present | present | |
| Non-grid slides (~1.0 mL) | may be present | present | |
| Cen-Slide | may be present | present | |
| Slide & Coverslip (~0.5 mL) | may be present | present | |
| Slide & Coverslip (~1.0 mL) | may be present | present | |
| FisherBrand UriSystem DeciSlide | may be present | present | |

| Analytes | Level 1 - 55251 | Level 2 - 55252 |
|--|--------------------------------------|---|
| 77 Elektronika (Visual / Analyzers) | | |
| Bilirubin | Negative | 1 - 6 mg/dL (1+ - 3+) ⁸ |
| Urobilinogen | Normal | 2 - 12 mg/dL (1+ - 4+) ⁸ |
| Ketones | Negative | 5 - 150 mg/dL ((+) - 3+) |
| Ascorbic Acid | Negative | Negative |
| Glucose | Normal | 50 - 1000 mg/dL (1+ - 4+) |
| Protein | Negative | 30 - 500 mg/dL (1+ - 3+) |
| Blood | Negative-10 Ery/µL (Neg - 1+) | 10 - 300 Ery/µL (1+ - 3+) |
| pH | 5 - 6 | 7 - 8 |
| Nitrite | Negative | Positive |
| Leukocytes | Negative | 75 - 500 Leu/µL (2+ - 3+) |
| Specific Gravity | 1.010 - 1.030 | 1.000 - 1.025 |
| Analyticon® Combi Screen (Visual) | | |
| Bilirubin | Negative | 1+ - 3+ |
| Urobilinogen | Normal | 2 - 12 mg/dL (35 - 200 µmol/l) ^{8*} |
| Ketones | Negative | (+) - 3+ |
| Ascorbic Acid | Negative | Negative |
| Glucose | Normal | 50 - 1000 mg/dL (2.8 - 56 mmol/L) |
| Protein | Negative | 30 - 500 mg/dL |
| Blood | Negative | 10 - 300 Ery/µL (1+ - 3+) |
| pH | 5 - 6 | 6 - 8 |
| Nitrite | Negative ⁹ | Positive |
| Leukocytes | Negative | 25 - 500 Leu/µL |
| Specific Gravity | 1.010 - 1.020 | 1.000 - 1.010 |
| Creatinine | 10 - 200 mg/dL (0.9 - 17.7 mmol/L) | 200 - 300 mg/dL (17.7 - 26.5 mmol/L) |
| Microalbumin | 10 - 80 mg/L | 150 - 500 mg/dL |
| Analyticon® (Analyzers) | | |
| Bilirubin | Negative | 1 - 4 mg/dL 17-70 µmol/L (1+ - 3+) |
| Urobilinogen | Normal | 2 - 12 mg/dL 35-200 µmol/L (1+ - 4+) ^{8*} |
| Ketones | Negative | 10 - 300 mg/dL 1.0 - 30 mmol/L (+) - 3+ |
| Ascorbic Acid | Negative - 20 mg/dL Negative - 1+ | Negative - 20 mg/dL Negative - 1+ |
| Glucose | Normal | 50 - 1000 mg/dL 2.8 - 5.6 mmol/L (1+ - 5+) |
| Protein | Negative | 30 - 500 mg/dL 0.3 - 5 g/L (1+ - 3+) |
| Blood | Negative | 10 - 300 Ery/µL (1+ - 3+) |
| pH | 5 - 7 | 6 - 9 |
| Nitrite | Negative ⁹ | Positive |
| Leukocytes | Negative | 25 - 500 Leu/µL (1+ - 3+) |
| Specific Gravity | 1.010 - 1.030 | 1.000 - 1.020 |

| Analytes | Level 1 - 55251 | Level 2 - 55252 |
|---|------------------------------|--|
| CYPRESS DIAGNOSTICS Urine Strips • CYANStrip • CYANStrip Mini • Visual | | |
| Urobilinogen | Normal (0.1-1mg/dL) | Normal-2 mg/dL (0.1-33 µmol/L) ¹¹ |
| Glucose | Negative | 50-2000 mg/dL (2.8-111 mmol/L) |
| Bilirubin | Negative | Small-Large (1+-3+) |
| Ketones | Negative | 5-40 mg/dL (0.5-4 mmol/L) (Trace-2+) |
| Specific Gravity | 1.015-1.025 | 1.005-1.020 |
| Blood | Negative-Trace | 10-250 RBC/µL (1+-3+) |
| pH | 5-6.5 | 7-9 |
| Protein | Negative | 15-300 mg/dL (0.15-3.0 g/L) (Trace-3+) |
| Nitrite | Negative | Positive |
| Leukocytes | Negative | 15-500 WBC/µL (Trace-3+) |
| Microalbumin | 10-50 mg/dL (0.5-0.9 mmol/L) | 100-300 mg/dL (8.8-26.5 mmol/L) |
| Creatinine | 10 mg/L | 30-150 mg/L |
| DiaLab Urine Strip Analyzer 500/Urine Strip 10C/Urine Strip 2MC | | |
| Leukocytes | Negative | 70-500 Leu/µL |
| Nitrite | Negative | Positive |
| Urobilinogen | 0.2 - 1 mg/dL | 0.2 - 4 mg/dL ¹¹ |
| Protein | Negative | 30 - 300 mg/dL |
| pH | 5.0 - 7.0 | 7.0 - 9.0 |
| Blood | Negative | 1+ - 3+ |
| Specific Gravity | 1.015 - 1.030 | 1.005 - 1.025 |
| Ketones | Negative | 5 - 160 mg/dL |
| Bilirubin | Negative | 1 - 4 mg/dL |
| Glucose | Negative | 100 - 500 mg/dL |
| Creatinine | 10 - 50 mg/dL | 50 - 200 mg/dL |
| Microalbumin | 1 - 3 mg/dL | 3 - 15 mg/dL |
| DFI CYBOW • ComboStik • DUS Urine Reagent Strips (Visual) | | |
| Urobilinogen | Normal (0.1-1 mg/dL) | Normal-2 mg/dL (0.1-33 µmol/L) ¹¹ |
| Glucose | Negative | 50-2000 mg/dL (2.8-111 mmol/L) |
| Bilirubin | Negative | Small-Large (1+-3+) |
| Ketones | Negative | 5-40 mg/dL (0.5-4 mmol/L) (Trace-2+) |
| Specific Gravity | 1.015-1.025 | 1.005-1.020 |
| Blood | Negative-Trace | 10-250 RBC/µL (1+-3+) |
| pH | 5-6.5 | 7-9 |
| Protein, Total | Negative | 15-300 mg/dL (0.15-3.0 g/L) (Trace-3+) |
| Nitrite | Negative | Positive |
| Leukocytes | Negative | 15-500 WBC/µL (Trace-3+) |
| Creatinine | 10-50 mg/dL (0.5-0.9 mmol/L) | 100-300 mg/dL (8.8-26.5 mmol/L) |
| Microalbumin | 10 mg/L | 30-150 mg/L |
| DFI CYBOW R-50 (50S) • ComboStik R-50 (50S) • DUS R-50 (50S) | | |
| Urobilinogen | Normal (0.1-1 mg/dL) | Normal-2 mg/dL (0.1-33 µmol/L) ¹¹ |
| Glucose | Negative | 50-2000 mg/dL (2.8-111 mmol/L) |
| Bilirubin | Negative | Small-Large (1+-3+) |
| Ketones | Negative | 5-40 mg/dL (0.5-4 mmol/L) (Trace-2+) |
| Specific Gravity | 1.015-1.025 | 1.005-1.020 |
| Blood | Negative-Trace | 10-250 RBC/µL (1+-3+) |
| pH | 5-6.5 | 7-9 |
| Protein, Total | Negative | 15-300 mg/dL (0.15-3.0 g/L) (Trace-3+) |
| Nitrite | Negative | Positive |
| Leukocytes | Negative | 15-500 WBC/µL (Trace-3+) |
| Creatinine | 10-50 mg/dL (0.5-0.9 mmol/L) | 100-300 mg/dL (8.8-26.5 mmol/L) |
| Microalbumin | 10 mg/L | 30-150 mg/L |
| DFI CYBOW Reader 300 • ComboStik R-300 • DUS R-300 | | |
| Urobilinogen | Normal (0.1-1 mg/dL) | Normal-2 mg/dL (0.1-33 µmol/L) ¹¹ |
| Glucose | Negative | 50-2000 mg/dL (2.8-111 mmol/L) |
| Bilirubin | Negative | Small-Large (1+-3+) |
| Ketones | Negative | 5-40 mg/dL (0.5-4 mmol/L) (Trace-2+) |
| Specific Gravity | 1.015-1.025 | 1.005-1.020 |
| Blood | Negative-Trace | 10-250RBC/µL (1+-3+) |
| pH | 5-6.5 | 7-9 |
| Protein, Total | Negative | 15-300mg/dL (0.15-3.0g/L) (Trace-3+) |
| Nitrite | Negative | Positive |
| Leukocytes | Negative | 15-500 WBC/µL (Trace-3+) |
| Creatinine | 10-50 mg/dL (0.5-0.9 mmol/L) | 100-300 mg/dL (8.8-26.5 mmol/L) |
| Microalbumin | 10 mg/L | 30-150 mg/L |

| Analytes | Level 1 - 55251 | Level 2 - 55252 |
|--|------------------------------|---|
| DFI CYBOW Reader 720 • Combostik R-700 • DUS R-720 • DFI CYBOW Reader 720 | | |
| Urobilinogen | Normal (0.1-1 mg/dL) | Normal-2 mg/dL (0.1-33 µmol/L) ¹¹ |
| Glucose | Negative | 50-2000 mg/dL (2.8-111 mmol/L) |
| Bilirubin | Negative | Small-Large (1+-3+) |
| Ketones | Negative | 5-40 mg/dL (0.5-4 mmol/L) (Trace-2+) |
| Specific Gravity | 1.015-1.025 | 1.005-1.020 |
| Blood | Negative-Trace | 10-250RBC/µL (1+-3+) |
| pH | 5-6.5 | 7-9 |
| Protein, Total | Negative | 15-300mg/dL (0.15-3.0g/L) (Trace-3+) |
| Nitrite | Negative | Positive |
| Leukocytes | Negative | 15-500 WBC/µL (Trace-3+) |
| Creatinine | 10-50 mg/dL (0.5-0.9 mmol/L) | 100-300 mg/dL (8.8-26.5 mmol/L) |
| Microalbumin | 10mg/L | 30-150 mg/L |
| ERBA LACHEMA DekAPHAN LAURA STRIPS & LAURA Urine Analyzer ERBA Mannheim Uro-dip 10e STRIPS & Uro-dipcheck 400e Urine Analyzer | | |
| Bilirubin | Negative | 3 - 6 mg/dL (51 - 103 µmol/L) (2+ - 3+) |
| Blood | Negative - Trace | 50 - 250 Ery/µL (2+ - 3+) |
| Glucose | Negative | 100 - 1000 mg/dL (5.5 - 55 mmol/L) (2+ - 4+) |
| Ketones | Negative | 52 - 156 mg/dL (5 - 15 mmol/L) (2+ - 3+) |
| Leukocytes | Negative | 75 - 500 Leu/µL (2+ - 3+) |
| Nitrite | Negative | Positive |
| pH | 5 - 6 | 6.5 - 8 |
| Protein | Negative | 30 - 500 mg/dL (0.3 - 5 g/L) (1+ - 3+) |
| Specific Gravity | 1,020 - 1,030 | 1,000 - 1,015 |
| Urobilinogen | Normal | 1 - 12 mg/dL (17 - 203 µmol/L) (1+ - 4+) ⁸ |
| ERBA LACHEMA DekAPHAN LAURA STRIPS & LAURA M Urine Analyzer ERBA Mannheim Uro-dip 10e STRIPS & LAURA M Urine Analyzer | | |
| Bilirubin | Negative | 3 - 6 mg/dL (51 - 103 µmol/L) (2+ - 3+) |
| Blood | Negative - Trace | 50 - 250 Ery/µL (2+ - 3+) |
| Glucose | Negative | 100 - 1000 mg/dL (5.5 - 55 mmol/L) (2+ - 4+) |
| Ketones | Negative | 16 - 156 mg/dL (1.5 - 15 mmol/L) (1+ - 3+) |
| Leukocytes | Negative | 75 - 500 Leu/µL (2+ - 3+) |
| Nitrite | Negative | Positive |
| pH | ≤6 | ≤6 - 8 |
| Protein | Negative | 30 - 500 mg/dL (0.3 - 5 g/L) (1+ - 3+) |
| Specific Gravity | 1,015 - 1,030 | 1,000 - 1,015 |
| Urobilinogen | Normal | 1 - 12 mg/dL (17 - 203 µmol/L) (1+ - 4+) ⁸ |
| ERBA LACHEMA DekAPHAN LAURA STRIPS & LAURA Smart Urine Analyzer ERBA Mannheim Uro-dip 10e STRIPS & Uro-dipcheck 240e Urine Analyzer | | |
| Bilirubin | Negative | 3 - 6 mg/dL (51 - 103 µmol/L) (2+ - 3+) |
| Blood | Negative - Trace | 50 - 250 Ery/µL (2+ - 3+) |
| Glucose | Negative | 100 - 1000 mg/dL (5.5 - 55 mmol/L) (2+ - 4+) |
| Ketones | Negative | 52 - 156 mg/dL (5 - 15 mmol/L) (2+ - 3+) |
| Leukocytes | Negative | 75 - 500 Leu/µL (2+ - 3+) |
| Nitrite | Negative | Positive |
| pH | 5 - 6.5 | 6.5 - 8 |
| Protein | Negative | 30 - 500 mg/dL (0.3 - 5 g/L) (1+ - 3+) |
| Specific Gravity | 1.020 - 1.030 | 1.000 - 1.015 |
| Urobilinogen | Normal | 1 - 12 mg/dL (17 - 203 µmol/L) (1+ - 4+) ⁸ |
| ERBA LACHEMA DekAPHAN LAURA STRIPS (Visual) ERBA Mannheim Uro-dip 10e STRIPS (Visual) | | |
| Bilirubin | Negative | 3 - 6 mg/dL (51 - 103 µmol/L) (2+ - 3+) |
| Blood | Negative - Trace | 50 - 250 Ery/µL (2+ - 3+) |
| Glucose | Negative | 300 - 1000 mg/dL (17 - 55 mmol/L) (3+ - 4+) |
| Ketones | Negative | 52 - 156 mg/dL (5 - 15 mmol/L) (2+ - 3+) |
| Leukocytes | Negative | 75 - 500 Leu/µL (2+ - 3+) |
| Nitrite | Negative | Positive |
| pH | 5 - 6.5 | 6.5 - 8 |
| Protein | Negative | 30 - 500 mg/dL (0.3 - 5 g/L) (1+ - 3+) |
| Specific Gravity | 1.025 - 1.030 | 1.000 - 1.015 |
| Urobilinogen | Normal | 1 - 12 mg/dL (17 - 203 µmol/L) (1+ - 4+) ⁸ |

| Analytes | Level 1 - 55251 | Level 2 - 55252 |
|---|-----------------------|---------------------------------------|
| Mindray UA-5800/UA-6800 | | |
| DATA NOT AVAILABLE AT THE TIME OF PRINTING. | | |
| ROCHE VISUAL TESTING (Visual Test Strips Only) | | |
| Specific Gravity | 1.015 - 1.030 | 1.000 - 1.010 |
| pH | 5 - 6 | 7 - 9 |
| Leukocytes | Negative | 75 - 500 Leu/ μ L (2+ - 3+) |
| Nitrite | Negative | Positive |
| Protein | Negative | 30 - 100 mg/dL |
| Glucose | Normal | 300 - 1000 mg/dL (3+ - 4+) |
| Ketones | Negative | 10 - 150 mg/dL (1+ - 3+) |
| Urobilinogen ^{9*} | Normal | 1 - 8 mg/dL (1+ - 3+) |
| Bilirubin ^{9*} | Negative | 1+ - 3+ |
| Blood | Negative - Trace | 10 - 250 Ery/ μ L (1+ - 4+) |
| Microalbumin ⁶ | Negative - 20 mg/L | 50 - 100 mg/L |
| ROCHE cobas 6500 (cobas u 601) | | |
| Blood | Negative - Trace | 150 - 250 Ery/ μ L |
| Leukocytes | Negative | 100 - 500 Leu/ μ L |
| Nitrite | Negative | Positive |
| Ketones | Negative | 15 - 150 mg/dL |
| Glucose | Normal | 300 - 1000 mg/dL |
| Protein | Negative ⁶ | 75 - 150 mg/dL ⁸ |
| Urobilinogen | Normal | 1 - 8 mg/dL ^{9*} |
| Bilirubin | Negative | 3 - 6 mg/dL * |
| pH | 5.0 - 6.0 | 7 - 9 |
| Specific Gravity | 1.019 - 1.030 | 1.011 - 1.022 |
| ROCHE Miditron M | | |
| Blood | Negative - Trace | 150 - 250 Ery/ μ L |
| Bilirubin | Negative | 3 - 6 mg/dL * |
| Urobilinogen | Normal | 1 - 8 mg/dL ^{9*} |
| Ketones | Negative | 15 - 150 mg/dL |
| Glucose | Normal | 300 - 1000 mg/dL |
| Protein | Negative ⁶ | 25 - 150 mg/dL ⁸ |
| Nitrite | Negative | Positive |
| Leukocytes | Negative | 100 - 500 Leu/ μ L |
| pH | 5 - 6.5 | 7 - 9 |
| Specific Gravity | 1.015 - 1.025 | 1.000 - 1.015 |
| ROCHE Urisys 1100 Urine Analyzer or ROCHE Urilux S Urine Analyzer | | |
| Blood | Negative - Trace | 50 - 250 Ery/ μ L (3+ - 4+) |
| Bilirubin | Negative | 3 - 6 mg/dL (2+ - 3+)* |
| Urobilinogen | Negative - Normal | 1 - 8 mg/dL (1+ - 3+)* |
| Ketones | Negative | 5 - 150 mg/dL ((+) - 3+) |
| Glucose | Negative - Normal | 300 - 1000 mg/dL (3+ - 4+) |
| Protein | Negative ⁶ | 25 - 150 mg/dL (1+ - 3+) ⁷ |
| Nitrite | Negative | Positive (1+) |
| Leukocytes | Negative | 100 - 500 Leu/ μ L (2+ - 3+) |
| pH | 5 - 6.5 | 7 - 9 |
| Specific Gravity | 1.015 - 1.025 | 1.000 - 1.015 |
| ROCHE Urisys 1800 Urine Analyzer or ROCHE Cobas u 411 Urine Analyzer | | |
| Blood | Negative - Trace | 150 - 250 Ery/ μ L (4+ - 5+) |
| Bilirubin | Negative | 3 - 6 mg/dL (2+ - 3+)* |
| Urobilinogen | Negative - Normal | 1 - 8 mg/dL (1+ - 3+)* |
| Ketones | Negative | 15 - 150 mg/dL (1+ - 4+) |
| Glucose | Negative - Normal | 300 - 1000 mg/dL (3+ - 4+) |
| Protein | Negative ⁶ | 75 - 150 mg/dL (2+ - 3+) ⁸ |
| Nitrite | Negative | Positive |
| Leukocytes | Negative | 100 - 500 Leu/ μ L (2+ - 3+) |
| pH | 5 - 6.5 | 7 - 9 |
| Specific Gravity | 1.015 - 1.025 | 1.000 - 1.020 |
| ROCHE Urisys 2400 | | |
| Blood | Negative - Trace | 150 - 250 Ery/ μ L |
| Bilirubin | Negative | 3 - 6 mg/dL |
| Urobilinogen | Normal | 1 - 8 mg/dL ^{9*} |
| Ketones | Negative | 15 - 150 mg/dL |
| Glucose | Normal | 300 - 1000 mg/dL |
| Protein | Negative ⁷ | 25 - 150 mg/dL ⁸ |
| Nitrite | Negative | Positive |
| Leukocytes | Negative | 100 - 500 Leu/ μ L |
| pH | 5 - 6.5 | 7 - 9 |
| Specific Gravity | 1.018 - 1.029 | 1.010 - 1.021 |
| YD Diagnostics URISCAN Urine Test Strips (Visual) | | |
| Blood | Negative | 10-250 RBC/uL (1+~3+) |
| Bilirubin | Negative | 0.5-3.0 mg/dL (1+~3+) ⁷ |
| Urobilinogen | Normal | 1-12 mg/dL (1+~4+) |
| Ketones | Negative | 10-100 mg/dL (1+~3+) |
| Protein | Negative | 30-1000 mg/dL (1+~4+) |
| Nitrite | Negative | Positive |
| Glucose | Negative | 250-2000 mg/dL (1+~4+) |
| pH | 5.0-6.5 | 7.0-9.0 |
| Specific gravity | 1.015-1.030 | 1.005-1.020 |
| Leucocytes | Negative | 25-500 WBC/uL (1+~3+) |

| Analytes | Level 1 - 55251 | Level 2 - 55252 |
|---|-----------------|------------------------------------|
| YD Diagnostics URISCAN Pro II • URISCAN Optima II • URISCAN Pro • URISCAN Optima | | |
| Blood | Negative | 10-250 RBC/uL (1+~3+) |
| Bilirubin | Negative | 0.5-3.0 mg/dL (1+~3+) ⁷ |
| Urobilinogen | Normal | 1-12 mg/dL (1+~4+) |
| Ketones | Negative | 10-100 mg/dL (1+~3+) |
| Protein | Negative | 30-1000 mg/dL (1+~4+) |
| Nitrite | Negative | Positive |
| Glucose | Negative | 250-2000 mg/dL (1+~4+) |
| pH | 5.0-6.5 | 6.5-8.5 |
| Specific gravity | 1.010-1.030 | 1.005-1.020 |
| Leucocytes | Negative | 25-500 WBC/uL (1+~3+) |
| Microalbumin | Negative-30mg/L | 30-150mg/L |
| Creatinine | 10-100mg/dL | 100-300mg/dL |
| YD Diagnostics URISCAN Super YD Diagnostics URISCAN Super+ | | |
| Blood | Negative | 10-250 RBC/uL (1+~3+) |
| Bilirubin | Negative | 0.5-3.0 mg/dL (1+~3+) ⁷ |
| Urobilinogen | Normal | 1-12 mg/dL (1+~4+) |
| Ketones | Negative | 10-100 mg/dL (1+~3+) |
| Protein | Negative | 30-1000 mg/dL (1+~4+) |
| Nitrite | Negative | Positive |
| Glucose | Negative | 250-2000 mg/dL (1+~4+) |
| pH | 5.0-6.5 | 6.5-8.5 |
| Specific gravity | 1.020-1.030 | 1.012-1.022 |
| Leucocytes | Negative | 25-500 WBC/uL (1+~3+) |
| YD URISCAN Super + | | |
| Blood | Negative | 10-250 RBC/uL (1+~3+) |
| Bilirubin | Negative | 0.5-3.0 mg/dL (1+~3+) ⁷ |
| Urobilinogen | Normal | 1-12 mg/dL (1+~4+) |
| Ketones | Negative | 10-100 mg/dL (1+~3+) |
| Protein | Negative | 30-1000 mg/dL (1+~4+) |
| Nitrite | Negative | Positive |
| Glucose | Negative | 100-1000 mg/dL (\pm -3+) |
| pH | 5.0-6.5 | 6.5-8.5 |
| Specific gravity | 1.020-1.030 | 1.012-1.022 |
| Leucocytes | Negative | 25-500 WBC/uL (1+~3+) |

Footnotes for values, Fußnoten für werte, Apostilles pour des valeurs, Note a piè di pagina per i valori, Notas al pie de la página para los valores

ENGLISH
1 Values only apply to Clinitek Microalbumin Reagent Strips when read on the Clinitek 50 and Status
2 Values only apply to Multistix Pro™ Reagent Strips
3 Values only apply to Multistix Pro and Clinitek Microalbumin Reagent Strips when read on Clinitek Urine Analyzers
4 VISUAL: Some customers may obtain false negatives.
5 Values apply to Chemstrip® Micral Reagent Strips
6 Some customers may obtain false positives.
7 Some customers may obtain false negatives.
8 Atypical color
9 Values only apply to Siemens Clinitek 50, 500
10 Absorbance at 620 nm
11 The urobilinogen reaction produces an atypical color which may result in a normal (0.2 E.U./dL) reading. Should this occur, a visual observation of the intensification of the pad color indicates a positive response.
• See Limitations













DEUTSCH
1 Werte gelten nur für Clinitek Mikroalbumin-Reagenzstreifen wenn diese auf Clinitek 50 und Status
2 Werte gelten nur für Multistix Pro™ Reagenzstreifen
3 Werte gelten nur für Multistix Pro und Clinitek Urin-Analysatoren
4 VISUAL: Manche Kunden erhalten möglicherweise falsch negative Ergebnisse.
5 Werte gelten für Chemstrip® Micral Reagenzstreifen
6 Manche Kunden erhalten möglicherweise falsch positive Ergebnisse.
7 Manche Kunden erhalten möglicherweise falsch negative Ergebnisse.
8 Atypische Farbe
9 Werte gelten nur für Siemens Clinitek 50, 500
10 Absorption bei 620 nm
11 Die Urobilinogen-Reaktion erzeugt eine atypische Farbe, die zu einem normalen Messwert (0.2 E.U/dL) führen kann. In diesem Fall kann eine positive Reaktion anhand der sichtbar veränderten Farbintensität des Testfeldes festgestellt werden.
• Siehe Einschränkungen


FRANÇAIS
1 Valeurs s'appliquent uniquement aux bandes de réactif Clinitek micro-albumine lues sur Clinitek 50 et Status
2 Valeurs s'appliquent uniquement aux bandes de réactif Multistix Pro™
3 Valeurs s'appliquent uniquement aux bandes de réactif Multistix Pro et Clinitek micro-albumine lues sur Clinitek Analyseurs d'urine
4 VISUAL: Certains clients sont susceptibles d'obtenir des faux négatifs.
5 Valeurs s'appliquent aux bandes de réactif Chemstrip® Micral
6 Certains clients sont susceptibles d'obtenir des faux positifs.
7 Certains clients sont susceptibles d'obtenir des faux négatifs.
8 Couleur atypique
9 Valeurs s'appliquent uniquement aux Siemens Clinitek 50, 500

10 Absorbance à 620 nm
11 La réaction de l'urobilinogène produit une couleur atypique pouvant donner lieu à une lecture normale (0.2 unité Ehrlich/dL). Si cela se produit, l'observation visuelle de l'intensification de la couleur de la zone de test indique une réponse positive.
• Voir Limitations


ITALIANO
1 I valori si riferiscono esclusivamente alle Strisce reagenti per microalbumina Clinitek lette su Clinitek 50 e Status
2 I valori si riferiscono esclusivamente alle Strisce reagenti Multistix Pro™
3 I valori si riferiscono esclusivamente alle Strisce reagenti per microalbumina Multistix Pro e Clinitek lette su Clinitek Analizzatori urine
4 VISUAL: Alcuni pazienti possono ottenere risultati falsi negativi.
5 I valori si riferiscono alle Strisce reagenti Micral Chemstrip®
6 Alcuni pazienti possono ottenere risultati falsi positivi.
7 Alcuni pazienti possono ottenere risultati falsi negativi.
8 Colore atipico
9 I valori si riferiscono esclusivamente alle Siemens Clinitek 50, 500
10 Assorbanza a 620 nm
11 La reazione dell'urobilinogeno produce un colore atipico che può determinare una lettura normale (0.2 U.E./dl). In questo caso, se si nota visivamente un'intensificazione del colore del cuscinetto, questo indica una reazione positiva.
• Vedere limiti

ESPAÑOL
1 Los valores son aplicables únicamente a las tiras reactivas Clinitek Microalbumin cuando se leen en equipos Clinitek 50 y Status
2 Los valores son aplicables únicamente a las tiras reactivas Multistix Pro™
3 Los valores son aplicables únicamente a las tiras reactivas Multistix Pro y Clinitek Microalbumin cuando se leen en equipos Clinitek Analizadores de orina
4 VISUAL: Algunos pacientes pueden obtener resultados negativos falsos.
5 Los valores son aplicables a las tiras reactivas Chemstrip® Micral
6 Algunos pacientes pueden obtener resultados positivos falsos
7 Algunos pacientes pueden obtener resultados negativos falsos.
8 Color anormal
9 Los valores son aplicables únicamente a las Siemens Clinitek 50, 500
10 Absorbancia a 620 nm
11 La reacción del urobilinógeno genera un color atípico que puede dar lugar a una lectura normal (0.2 E.U./dl). Si ocurriera esto, la respuesta es positiva si se observa visualmente una intensificación del color de la almohadilla.
• Ver las limitaciones


| | |
|---|--|
|  | European Conformity CE-Konformitätskennzeichnung Conformité aux normes européennes Conformità europea Conformidad europea |
|  | Catalog No. Bestellnr. N° de catalogue Catalogo n. N° de catálogo |
|  | Manufactured by Hergestellt von Fabriqué par Fabbricato da Fabricado por |
|  | Authorized Representative Bevollmächtigter Représentant agréé Rappresentante autorizzato Representante autorizado |
|  | Consult instructions for use Gebrauchsanweisung beachten Consulter les instructions d'utilisation Consultare le istruzioni l'uso Consulte las instrucciones de uso |
|  | Lot Number Bezeichnung Désignation du lot Numero di lotto Denominación de lote |
|  | Caution, See Product Insert Achtung, Siehe Packungsbeilage Attention, voir notice d'utilisation Attenzione, vedere il foglio illustrativo del prodotto Atención, consulte el folleto del producto |
|  | For in vitro diagnostic use In-vitro Diagnosticum Pour diagnostic in vitro Per uso diagnostico in vitro De uso diagnóstico in vitro |
|  | Temperature limitation Temperaturbegrenzungen Limites de température limiti di temperatura límite de temperatura |
|  | Contents of kit Inhalt der Packung Contenu du coffret Contenuto della confezione Contenido del estuche |
|  | Use by (last day of month) Verwendbar bis (letzter Tag des Monats) Utilisable jusqu'à (dernier jour du mois indiqué) Da utilizzare prima del (ultimo giorno del mese) Estable hasta (ultimo día del mes) |
|  | Biological Risk Biogefährdung Risque biologique Rischio biologico Peligro biológico |




Level 1 **LOT** 55251



Level 2 **LOT** 55252



Level 1&2
Expiration Date

 2022-08-31




2005 Manhattan Beach Blvd.
Redondo Beach, CA 90278-1205 USA
phone: (310) 536-0006
fax: (310) 536-9977
www.quantimetrix.com

E-M047025A-10/20



Quantimetrix Corporation
2005 Manhattan Beach Blvd.
Redondo Beach, CA 90278
+1.310.536.0006



MDSS GmbH
Schiffgraben 41
30175 Hannover, Germany