

Technical Update: Dipper® Urinalysis Dipstick Control

Lots #44551

Tech number: 157-18-44-1

Users: All Dipper Urinalysis Dipstick Control Users • Date issued: 06/06/18



English

Endogenous crystalline sediment may be present and does not affect the performance of the product.

This area is intentionally left blank.